

CL60

Electrolyte Analyzer

A Range Of Automated (ISE) Electrolyte Analyzer



7" color touch screen , easy to operate

High Accurate and long life electrode

TCO @ Sensor programmable multi –format print out.

Reagent pack ,real time monitoring of reagent residual volume

Up to 50,000 test results can be stored

RS -232 port, supporting bar code reader and USB interface.

Sleep mode to reduce reagent consumption

Option: Auto Sampler.



CHEMISCENT
BY JITM C GENES

| TECHNICAL SPECIFICATIONS | |
|---------------------------|---------------------------------------------------------------|
| METHODS OF CALCULATIONS | Ion Selective Electrode (ISE) Method |
| SAMPLER METHOD | Manual |
| SAMPLER VOLUME | Adjustable 100µl to 200µl |
| TEST WAVELENGTH | 470nm, 540nm 620nm |
| COMMUNICATION INTERFACE | Communication Interface |
| PRINTER | Inbuilt-Thermal paper Size 58mm and have option of Bar Reader |
| QUALITY CONTROL | Three Levels of quality can be entered (Low, Middle, High) |
| MEASURING PARAMETER | Sodium, Potassium, Chloride, Calcium, pH , Magnesium |
| FULL TIME | 45 Second |
| DATA STORAGE | 2000 Samples |
| DISPLAY | 320x240 LCD, Touch Screen |
| MEASURING TIME | 30 Second |
| POWER/DIMENSIONS | |
| Voltage | 110/220V |
| Frequency | 50/60Hz |
| Maximum Power | <50W |
| Operating Environment | 10°C to 32°C |
| Dimensions | 210mm x 340mm x 380mm |
| Maximum Relative Humidity | 85% |
| Weight | 5kgs |



JITM C Genes Pvt. Ltd.

H209, 1st Floor, Sector-63, Noida, U.P.-201301 (INDIA)
info@jitmcgenes.com, www.jitmcgenes.com | Customer Care No.: 87458 05386

ISO 13485:2016 Certified Company

